



8

D

С

 $\rightarrow$ 

В

Α

| PRODUCT   | 3.5 SINGLE-SIDED INNORACK IVC XL TALL RAT  |
|---|--|
| POWER SUPPLY ELECTRICAL INPUT<br>SPECIFICATIONS | VOLTAGE RANGE <sup>(1)</sup> : 85-264VAC<br>FREQUENCY RANGE: 47-63HZ<br>AC CURRENT(TYP.): 7.6A/115VAC, 3.6A/230VAC |
| OPERATIONAL AC CURRENT(TYP.) <sup>(2)</sup>     | .6-2.5A/120VAC, .3-1.25A/240VAC  |
| PLUG TYPE                                       | NEMA 5-15 PLUG/US, CEE 7/7 PLUG/EURO AND IEC 320<br>CONNECTOR  |
| LENGTH OF CORD                                  | US: 108 IN. [274 CM]<br>EUROPE: 78.7IN [200CM]   |

7

6

| PRODUCT CODE   | COLUMNS | ROWS | CAGE POSITIONS | LENGTH (L)  |                  | HEIGHT WITHOUT BLOWERS (H) |                  | HEIGHT WITH BLOWERS (H*) |                  | WEIGHT WITHOUT CAGES |                | HVAC <sup>(3)</sup>         |
|----------------|---------|------|----------------|-------------|------------------|----------------------------|------------------|--------------------------|------------------|----------------------|----------------|-----------------------------|
|                |         |      |                | inches (in) | centimeters (cm) | inches (in)                | centimeters (cm) | inches (in)              | centimeters (cm) | pounds (lbs)         | kilograms (kg) | cubic feet per min<br>(CFM) |
| RS35T-2-6-12XL | 2       | 6    | 12             | 62.66       | 159.1            | 74.50                      | 189.2            | 89.25                    | 226.7            | 203                  | 92.3           | 83                          |
|                |         |      |                |             |                  |                            |                  |                          |                  |                      |                |                             |

 $\square$ 

4

5

5

4

3

- (1) DERATING MAY BE NEEDED UNDER LOW INPUT VOLTAGES.
- (2) VARIES DEPENDING ON RACK SIZE, ACH AND CONDITION OF BLOWER FILTERS COMPONENTS.
- (3) MEASUREMENTS LISTED ARE +/- 0.5".
- (4) TO ENSURE BEST RESULTS A THIMBLE OR ADJUSTABLE VALVE SHOULD BE INSTALLED AT EACH CEILING DROP
  HVAC VALUES ARE BASED ON RACK RUNNING AT MAXIMUM NEGATIVE AIRFLOW SETTING.
- (5) PLEASE CONTACT INNOVIVE FOR MORE INFORMATION.

THIS DOCUMENT CONTAINS PROPRIETARY INFORMATION WHICH MAY NOT BE DISCLOSED TO OTHERS, REPRODUCED OR USED WITHOUT AUTHORITY FROM INNOVIVE. SPECIFICATIONS SUBJECT TO CHANGE WITHOUT NOTICE.

D

С

В

А

## Innovive

3.5 SINGLE SIDED XL TALL RAT

2

3